



WRP Asia Pacific Sdn Bhd

1 4 7 8 1 7 V

Lot 1, Jalan 3, Kawasan Perusahaan
Bandar Baru Salak Tinggi,
43900 Sepang,
Selangor Darul Ehsan, MALAYSIA

Office +60-3-8706 1486

Facsimile +60-3-8706 1488

Website www.wrpworld.com

Ref No.: WRP/AP/2019/0931

EC DECLARATION OF CONFORMITY

We, **WRP Asia Pacific Sdn Bhd**, being the manufacturer for the medical devices as described hereafter:

**DERMAGRIP SafeDon Nitrile Examination Gloves Ultra LT Latex Free Powder Free
Non Sterile**

Size XS: D1500-79

Size S: D1501-79

Size M: D1502-79


Size L: D1503-79

Size XL: D1504-79

declare under our own responsibility that the above product in Class I medical devices as per Rule 5 of Annex IX is manufactured in conformity with the procedure relating to the EC declaration of conformity set out in Annex VII and meets the essential requirements of Council Directive 93/42/EEC as amended by 2007/47/EC which apply to them under the supervision of the notified body British Standards Institution, and carrying an identification number of 2797.

The product is in conformity with the provisions of Regulation (EU) 2016/425 and, where such is the case, with the national standard transposing harmonized standard No. EN 420:2003 +A1:2009, EN ISO 374-1:2016, EN 374-2:2014, EN 16523-1:2015, EN 374-4:2013 and EN ISO 374-5: 2016 is subject to the procedure set out in Module D of Regulation (EU) 2016/425 is identical to the PPE which is the subject of EC certificate of conformity No. CE 688314 issued by BSI (2797), Say Building, John M. Keynesplein 9, 1066 EP, Amsterdam, Netherlands.

Done at WRP Asia Pacific Sdn Bhd, on 19th April 2019.



Dato' Lee Son Hong
Chief Executive Officer/CEO
WRP Asia Pacific Sdn Bhd

Representative Office in the EEA
REMESCO Handelsges.m.b.H
Grinzinger Allee 5/25
A-1190 Vienna
Austria